



SCOPE OF ACCREDITATION TO ISO/IEC 17025:2017
& ANSI/NCSL Z540-1-1994

J&W INSTRUMENTS
 4800 Mustang Circle
 New Brighton, MN 55112
 Diane Naegeli Phone: 763 784 5708

CALIBRATION

Valid To: July 31, 2023

Certificate Number: 1752.01

In recognition of the successful completion of the A2LA evaluation process, accreditation is granted to this laboratory to perform the following calibrations^{1,5}:

I. Electrical – DC/Low Frequency

Parameter/Equipment	Range	CMC ^{2,4} (±)	Comments
DC Voltage – Measure	(0 to 100) mV (0 to 100) V	20 $\mu\text{V}/\text{V} + 0.3 \mu\text{V}$ 90 $\mu\text{V}/\text{V} + 30 \mu\text{V}$	Agilent 3458A
	(0 to 100) mV	0.27 mV + 0.02 mV	TRX-II process calibrator, Beamex calibrator
DC Current – Measure	(0 to 10) mA (10 to 100) mA	36 $\mu\text{A}/\text{A} + 0.05 \mu\text{A}$ 58 $\mu\text{A}/\text{A} + 0.05 \mu\text{A}$	Agilent 3458A
	(0 to 12) mA (12 to 24) mA	0.03 % + 0.01 mA 0.03 % + 0.01 mA	TRX-II process calibrator, Beamex calibrator
AC Current – Measure	(0 to 40) A	1.3 A	Fluke 33 clamp on ammeter
Resistance – Measure	(0 to 10) Ω (10 to 1000) Ω (1 to 10) k Ω	30 $\mu\Omega/\Omega + 50 \mu\Omega$ 17 $\mu\Omega/\Omega + 0.5 \mu\Omega$ 17 $\mu\Omega/\Omega + 5 \mu\Omega$	Agilent 3458A
	(0 to 400) Ω 400 Ω to 2 k Ω	1.5 % + 0.09 Ω 0.3 % + 0.4 Ω	TRX-II process calibrator, Beamex calibrator

Parameter/Equipment	Range	CMC ² (±)	Comments
Thermocouple Simulation –			
Type B	(50 to 1820) °C	0.95 °C	TRX-II process calibrator, Beamex calibrator
Type J	(-210 to 1200) °C	0.4 °C	
Type K	(-270 to 1370) °C	0.4 °C	
Type N	(-270 to 1300) °C	0.4 °C	
Type R	(-50 to 1769) °C	0.88 °C	
Type S	(-50 to 1769) °C	0.89 °C	
Type T	(-270 to 400) °C	0.28 °C	
RTD Simulation ³ –			
Pt 385, 100 Ω	(-200 to 850) °C	0.4 °C	TRX-II process calibrator, Beamex calibrator
Pt 392, 100 Ω	(-200 to 510) °C	0.4 °C	

II. Mechanical

Parameter/Equipment	Range	CMC ² (±)	Comments
Pressure Gauges ³	(0 to 300) psi	0.41 psi	DPI-603
	(0 to 1000) psi	0.61 psi	3D digital gauge
Vacuum Gauges ³	(0 to 27) inHg	0.44 inHg	DPI-601
	(-30.5 to 610.8) inHg	0.2 inHg	DPI-610

III. Thermodynamics

Parameter/Equipment	Range	CMC ^{2,6} (±)	Comments
Temperature – Measure	(192 to 660) °C	0.03 °C	Fluke 5628 PRT
	(-50 to 300) °C	0.4 °C	RTD, TRX-II, Beamex
	(300 to 1200) °C	0.93 °C + 0.4 % of reading	T/C, TRX-II, Beamex
Temperature – Measuring Equipment, Fixed Points	0 °C	0.04 °C	ICE point, Fluke 5628 PRT
	Ambient to 200 °C	0.3 °C	PRT, Beamex
Infrared – Measuring Equipment	(50 to 400) °C	2.5 °C	BR400 Black Body/Fluke 5628 PRT

IV. Time & Frequency

Parameter/Equipment	Range	CMC ² (±)	Comments
Stopwatches and Timing Devices ³	(0 to 24) hours	1 s	Stopwatch

¹ This laboratory offers commercial calibration service and field calibration service, where noted.

² Calibration and Measurement Capability Uncertainty (CMC) is the smallest uncertainty of measurement that a laboratory can achieve within its scope of accreditation when performing more or less routine calibrations of nearly ideal measurement standards or nearly ideal measuring equipment. CMCs represent expanded uncertainties expressed at approximately the 95 % level of confidence, usually using a coverage factor of $k = 2$. The actual measurement uncertainty of a specific calibration performed by the laboratory may be greater than the CMC due to the behavior of the customer's device and to influences from the circumstances of the specific calibration.

³ Field calibration service is available for this calibration. Please note the actual measurement uncertainties achievable on a customer's site can normally be expected to be larger than the CMC found on the A2LA Scope. Allowance must be made for aspects such as the environment at the place of calibration and for other possible adverse effects such as those caused by transportation of the calibration equipment. The usual allowance for the actual uncertainty introduced by the item being calibrated, (e.g. resolution) must also be considered and this, on its own, could result in the actual measurement uncertainty achievable on a customer's site being larger than the CMC.

- ⁴ The stated measured values are determined using the indicated instrument (see Comments). This capability is suitable for the calibration of the devices intended to measure or generate the measured value in the ranges indicated. CMC's are expressed as either a specific value that covers the full range or as a percent or fraction of the reading plus a fixed floor specification.
- ⁵ This scope meets A2LA's *P112 Flexible Scope Policy*.
- ⁶ The type of instrument or material being calibrated is defined by the parameter. This indicates the laboratory is capable of calibrating instruments that measure or generate the values in the ranges indicated for the listed measurement parameter.



Accredited Laboratory

A2LA has accredited

J&W INSTRUMENTS, INC.

New Brighton, MN

for technical competence in the field of

Calibration

This laboratory is accredited in accordance with the recognized International Standard ISO/IEC 17025:2017 *General requirements for the competence of testing and calibration laboratories*. This laboratory also meets the requirements of ANSI/NCSL Z540-1-1994 and R205 – Specific Requirements: Calibration Laboratory Accreditation Program. This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system (refer to joint ISO-ILAC-IAF Communiqué dated April 2017).



Presented this 8th day of July 2021.

A handwritten signature in blue ink, positioned above a horizontal line.

Vice President, Accreditation Services
For the Accreditation Council
Certificate Number 1752.01
Valid to July 31, 2023

For the calibrations to which this accreditation applies, please refer to the laboratory's Calibration Scope of Accreditation.